Radiation Therapy

Dr. Aastha Mishra

Introduction:-

Radiation therapy may be defined as treatment by means of radiations. The term Actino-therapy is also used for this.

Radiation therapy may include:

1. Infrared radiations
2. Ultraviolet radiations.

Infrared radiations

Introduction :-The infrared rays are electromagnetic waves with the wavelengths of 750 to 400000 nm

and frequency 4 × 1014 Hz and 7.5 × 1011 Hz. It lies beyond the red boundary of visible spectrum.

* Any hot body can produce infrared rays like the sun, electric bulb, coal fire, gas fire, etc. Sun is the natural source of infrared radiations.
* Infrared radiations can be produced by artificial generators.
* In the Physiotherapy departments infrared rays are produced by two

types of generators:

1. Nonluminous generators

2. Luminous generators.

Nonluminous generators provide infrared rays only whereas luminous generators emit infrared rays, visible as well as ultraviolet rays.

Therefore, nonluminous generators are termed as infrared radiation generators because they emit only infrared rays.

The heat produced by luminous generator is called the ‘radiant heat’.

Nonluminous Generators :-

Nonluminous generator consists of a simple type of element or coil wound on a cylinder of some insulating material such as fireclay or porcelain.

* An electric current is passed through the wire which results in the production of heat. This heat produces infrared rays which are transmitted through the porcelain.
* Porcelain gets heated by the method of conduction but the radiations generated in this way also include some of the visible rays.
* Therefore to avoid this, the coil is embedded in fireclay or porcelain or placed behind fireclay.

* Now the emission of rays is entirely from the fireclay which is commonly painted black and thus very few visible rays are produced.

* The element or the coil is thus placed at the focal point of a parabolic or spherical reflector.

* The reflector is mounted on a stand and its position can be adjusted as required

*All of the nonluminous generators take some time to get heated up for the production of infrared radiations, so they should be switched on before 5–7 minutes of the treatment.

Luminous Generators:-

Luminous generators emit infrared, visible and a few ultraviolet rays. These generators are in the form of incandescent lamps or bulbs.

* An incandescent lamp consists of a wire filament enclosed in a glass bulb, which may contain an inert gas at low pressure.
* The filament is a coil of fine wire which is usually made up of tungsten.
* Tungsten is a metal which is used because it can tolerate repeated heating and cooling.
*Incandescent bulb is usually mounted at the center of the parabolic reflector and the reflector is mounted on an adjustable stand.

* These luminous generators emit the electromagnetic waves with the wavelength in between 350 and 4000 nm.

* The front of the bulb is usually red so as to filter out the shorter visible and the ultraviolet rays.